

INSTRUCTIONS FOR USE




NewGen Surgical Needle Counters

Please read all information carefully.

Intended Use

The NewGen Surgical Needle Counter is a single-use, sterile device designed to safely facilitate correct count and disposal of needles and blades during surgical procedures.

www.newgensurgical.com/ifu



Warnings and Precautions

NewGen Surgical Needle Counters are manufactured with bio-based materials. With these materials, some discoloration, texture variation, or fiber inclusions may be present. Cosmetic appearance does not affect the performance, safety, nor sterility of the device.

To ensure proper functioning, keep dry. Do not use as a fluid reservoir. Do not submerge or expose the needle counter to water or other liquids.

To ensure secure closure, do not separate needle counter box.

When using needle counters with magnets, caution should be used during procedures with materials or equipment which may be adversely affected by magnetic fields.

ASTM F2132 compliant for puncture resistance. Do not use excessive force that could compromise the integrity of the needle counter box.

This device is manufactured and sterilized for single-use only.

Do not reuse, reprocess, or resterilize. Reuse, reprocessing, or resterilization may compromise the structural integrity of the device and/or lead to device failure.

Supports OSHA recommendations for safe sharps handling. Always refer to sharps protocol for your facility.







Store in a cool, dry place.








Do not use if product is damaged. Do not use if package is open or damaged.

Any serious incident that has occurred in relation to the product should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Instructions for Use

1. Remove tabs to expose adhesive on the front and back of the needle counter container.
2. Open needle counter and place on flat surface.
 - 2a. For foam block configurations, remove the adhesive backing, replace, and press firmly to secure foam.
3. Place used needles into container during procedure; one per numbered space.
4. Place larger needles and blades flat on magnetic surface.
5. Securely close needle counter once container is at capacity and/or procedure has ended.
6. Dispose the needle counter in compliance with facility guidelines for disposing of sharps.

	Ethylene oxide sterilizable
	See Instructions for Use
	Do not reuse
	Batch code
	Use by YYYY-MM-DD or YYYY-MM
	Manufacturer

	Catalog number
	Keep dry
	Do not use if package is opened or damaged
	Latex-free
	DEHP-free
	PVC-free
	European Conformity

 NGSNC20MA, NGSNC30MA, NGSNC40FMMA, NGSNC60MA, NGSNC80FFMMA
NGSNC100FMMA

US Patent D804053, D853582, D895839



NewGen Surgical, Inc.
San Rafael, CA 94901 USA
855.295.4500
www.newgensurgical.com

© 2023 NewGen Surgical, Inc.
Manufactured in China
45-0504-I

